

# MASTAZYME™ dsDNA

Enzymimmunoassay zur Bestimmung von Autoantikörpern gegen  
Doppelstrang-DNA (dsDNA) in Serum und Plasma

Enzyme immunoassay for the detection of autoantibodies to double-stranded DNA  
(dsDNA) in serum and plasma

## Gebrauchsinformation / Instructions for Use



**Nur zur *in-vitro* Diagnostik / For *in vitro* diagnostic use only**



Deutsch:            Seiten 03–08



English:            Pages 09–14

MASTAZYME™ dsDNA

REF 732020

12 x 8 Tests

**Lagerung / Storage: 2–8 °C**

**Erläuterungen zu den auf den Etiketten verwendeten graphischen Symbolen:**

**Explanations of abbreviations and icons used on labels:**

<b>STRIPS</b>		Mikrotiterstreifen	Microtiter strips
<b>CONTR</b>	<b>+</b>	Positive Kontrolle	Positive control
<b>CONTR</b>	<b>-</b>	Negative Kontrolle	Negative control
<b>CAL</b>	<b>1-4</b>	Kalibrator 1-4	Calibrator 1-4
<b>CONJ</b>		HRP-Konjugat	HRP conjugate
<b>DIL</b>		Probenverdünnungspuffer	Sample diluent
<b>SUBS</b>		TMB-Substrat	TMB substrate
<b>STOP</b>		Stopp-Lösung	Stopping solution
<b>WASH</b>	<b>CONC</b>	Waschpuffer, konzentriert	Washing buffer, concentrate
<b>LOT</b>		Charge	Batch
		Gebrauchsinformation beachten	Read instructions for use
		Wichtige Hinweise beachten	Important notes
		Verfallsdatum	Expiry Date
		Lagerung bei	Storage
		Reagenzien nicht zur Wiederverwendung geeignet	Reagents not reusable

	<b>Inhalt</b>	<b>Seite</b>
1.	Verwendungszweck	4
2.	Testprinzip	4
3.	Packungsinhalt	4
4.	Zusätzlich benötigte Materialien	5
5.	Warnhinweise und Vorsichtsmaßnahmen	5
6.	Lagerung und Stabilität	6
7.	Probengewinnung und -handhabung	6
8.	Testdurchführung	6
9.	Auswertung und Interpretation	7
10.	Testcharakteristika	8
11.	Literatur	8

## 1. Verwendungszweck

Der MASTAZYME™ dsDNA ELISA dient dem Nachweis von Autoantikörpern gegen die entsprechenden Autoantigene in Serum oder Plasma. MASTAZYME™ dsDNA ist nur zur *in-vitro* Diagnostik zu verwenden.

## 2. Testprinzip

Das Testprinzip der ELISAs kann in 4 Schritten beschrieben werden.

### 2.1 Seruminkubation und 1. Waschschrift

Spezifische Antikörper bilden mit Antigen, das an die Festphase gebunden ist, einen stabilen Immunkomplex. Nach einer 30-minütigen Inkubation bei Raumtemperatur werden unspezifisch gebundene Serumkomponenten durch Waschen entfernt.

### 2.2 Konjugatinkubation und 2. Waschschrift

Meerrettich-Peroxidase-markiertes Anti-Human-IgG bindet an die entsprechenden Antikörper auf dem Festphasenantigen und bildet mit diesen einen stabilen Immunkomplex. Überschüssiges nicht gebundenes Konjugat wird nach der 30-minütigen Inkubation durch Waschen entfernt.

### 2.3 Substrat- und Stoppreaktion

Nach Zugabe des TMB-Substrats wird dieses durch das Enzymkonjugat umgesetzt. Es entsteht eine bläuliche Färbung, deren Intensität mit der Menge der gebundenen Konjugatmoleküle korreliert. Nach 15 Minuten Inkubation wird die Reaktion durch Zugabe von 0,5 N Schwefelsäure (H<sub>2</sub>SO<sub>4</sub>) gestoppt. Die pH-Verschiebung führt zum Farbumschlag von blau nach gelb.

### 2.4 Auswertung

Die Reaktionsansätze können nun mit einem ELISA-Plattenreader bei 450 nm (empfohlene Referenzwellenlänge bei bichromatischer Messung: 600–690 nm) gemessen werden. Die Extinktion (OD) korreliert mit der Konzentration der spezifischen Antikörper.

Das Ergebnis kann aus einer Eichkurve abgelesen oder durch geeignete elektronische Kurvenberechnung (4-Parameter-Anpassung, Spline-Approximation o.ä.) ermittelt werden.

## 3. Packungsinhalt

Das Testkit enthält genügend Reagenzien für 12 x 8 = 96 Bestimmungen. Die Streifen der Mikrotiterplatte sowie alle anderen Reagenzien sind bei 2–8 °C zu lagern. Das Verfallsdatum der einzelnen Komponenten ist auf den jeweiligen Etiketten vermerkt.

12	Streifen	mit je 8 einzeln abbrechbaren Wells, beschichtet mit dem dsDNA Antigen
1 x	Rahmen	für Streifen der Mikrotiterplatte (MTP)
4 x 1,5 mL	Kalibratoren 1–4	Stabilisiertes humanes Plasma, enthält antigenspezifische Antikörper, gebrauchsfertig

	Anti-dsDNA Konzentration [IU/mL] nach WO/80
Kalibrator 1	1
Kalibrator 2 (Cut-off)	25
Kalibrator 3	50
Kalibrator 4	200
Positive Kontrolle	100
Negative Kontrolle	15

1 x 1,5 mL	Negative Kontrolle	Stabilisiertes humanes Plasma, gebrauchsfertig, kann in die Kurve integriert werden
1 x 1,5 mL	Positive Kontrolle	Stabilisiertes humanes Plasma, gebrauchsfertig, kann in die Kurve integriert werden
2 x 60 mL	Probenverdünnungspuffer	PBS Puffer mit Konservierungsmittel, gebrauchsfertig
1 x 12 mL	Enzymkonjugat	Meerrettich-Peroxidase markiertes Anti-Human-IgG (Ziege), gebrauchsfertig
1 x 12 mL	TMB-Substrat	3,3',5,5' Tetramethylbenzidin, gebrauchsfertig
1 x 12 mL	Stopplösung	0,5 N H <sub>2</sub> SO <sub>4</sub> (Schwefelsäure), gebrauchsfertig
2 x 50 mL	Waschpuffer 10 x Konz.	PBS/Tween Puffer, vor Gebrauch 1:10 mit dest. Wasser verdünnen, evtl. vor Gebrauch kurz auf 37 °C erwärmen, um mögliche Kristalle zu lösen

#### 4. Zusätzlich benötigte Materialien

- 5 µL-, 100 µL- und 500 µL-Pipetten oder Multipipetten
- Photometer für Mikrotiterplatten mit 450 nm Filter (Referenzfilter 600–690 nm)
- Waschgerät für Mikrotiterplatten (bei manuellem Waschen: Waschflasche)
- Röhrchen für Serumverdünnungen
- Messzylinder
- Destilliertes Wasser oder Wasser höherer Qualität

#### 5. Warnhinweise und Vorsichtsmaßnahmen

- Den Test nur zur *in-vitro* Diagnostik verwenden! Reagenzien nicht schlucken oder einatmen. Die Sicherheitsbestimmungen des Labors sind zu beachten. Im Labor darf nicht gegessen, getrunken oder geraucht werden.
- Alle Seren und Plasmen sowie Puffer des Kits, die humanes Probenmaterial enthalten, wurden mit anerkannten Methoden auf HBsAg, HIV und HCV getestet und für negativ befunden. Da das Vorhandensein solcher Erreger trotzdem nicht völlig ausgeschlossen werden kann, sollten die Reagenzien wie potenziell infektiöses Material behandelt werden.
- Serum- und Reagenzien-Kontaminationen sollten mit Desinfektionsmitteln gesäubert und der Abfall entsprechend entsorgt werden.
- Alle Reagenzien müssen vor Testbeginn auf Raumtemperatur (18–24 °C) gebracht werden.
- Vor der Verwendung sind die Reagenzien gut zu mischen. Heftiges Schütteln und Schaumbildung sind zu vermeiden.
- Beim Pipettieren ist auf gleiche Zeitintervalle zu achten, um für alle Testansätze gleiche Bedingungen zu gewährleisten.
- Beim Öffnen der Fläschchen ist eine Kontamination des Stopfens zu vermeiden. Um das Risiko möglicher Oxidationen zu minimieren, sind die Fläschchen nach Gebrauch sofort wieder zu verschließen.
- Um Verschleppungen und Kreuzkontaminationen zu vermeiden, sind Einmal-Pipettenspitzen zu verwenden.
- Reagenzien verschiedener Kit-Chargen sollten nicht verwendet werden.
- Alle Reagenzien sind vor Ablauf des Verfallsdatums zu verwenden.

- Gemäß den GLP (Good Laboratory Practice) oder entsprechenden Richtlinien sind alle Labogeräte regelmäßig auf Funktion und Präzision zu prüfen, dies gilt z. B. für die Pipetten, Waschgeräte und ELISA-Reader.
- Der Kontakt mit der Schwefelsäure enthaltenden Stopplösung und TMB-Substrat ist zu vermeiden. Bei Hautkontakt unverzüglich und gründlich mit Wasser abwaschen. Alle Geräte sofort nach Gebrauch sorgfältig reinigen.

## 6. Lagerung und Stabilität

Alle Reagenzien bei 2–8 °C lagern.

Das Verfallsdatum jedes Kitbestandteils ist auf dem entsprechenden Etikett vermerkt. Die Reagenzien nach Ablauf des Verfallsdatums nicht weiter verwenden.

Verdünnter Waschpuffer kann bei 2–8 °C gelagert werden. Unter diesen Bedingungen kann er bis zu 4 Wochen verwendet werden. Er muss allerdings vor Benutzung auf Raumtemperatur gebracht werden.

Nach dem Öffnen ist das angebochene Kit innerhalb von 3 Monaten zu verwenden.

## 7. Probengewinnung und -handhabung

Es kann sowohl Serum als auch Plasma (EDTA, Heparin) zur Bestimmung verwendet werden. Die Proben können 3 Tage bei 2–8 °C aufbewahrt werden. Bei längerer Aufbewahrung Proben sofort nach Entnahme in Aliquots aufteilen und bei -20 °C einfrieren.

Wiederholtes Einfrieren und Auftauen der Proben ist zu vermeiden!

Aufgetaute Proben sollten vor der Verwendung im Test gemischt (Vortex) werden.

Lipämische, ikterische oder hämolysierte Proben können falsche Ergebnisse ergeben.

Vor der Analyse müssen Patientenproben mit Probenverdünnungspuffer 1:101 (z. B. 5 µL Serum + 500 µL Probenverdünnungspuffer) verdünnt werden.

## 8. Testdurchführung

### 8.1. Vorbereitung der Reagenzien

Alle Reagenzien und Proben mischen und auf Raumtemperatur (RT, 18–24 °C) bringen.

**Waschpuffer:** Beim Vorliegen von Salzkristallen das Konzentrat auf 37 °C erwärmen und nach Lösung der Kristalle mischen.

Das Waschpuffer-Konzentrat mit destilliertem Wasser 1:10 verdünnen (z. B. 50 mL Konzentrat + 450 mL dest. Wasser), mischen.

- Die Gebrauchsanweisung ist zu befolgen. Jegliche Abänderung oder Modifikation erfolgt in Verantwortung des Anwenders.
- Alle Reagenzien müssen vor Testbeginn auf Raumtemperatur gebracht werden. Die Reagenzien sollten nur so lange wie nötig bei Raumtemperatur gelagert werden.
- Für die quantitative Auswertung ist mit jedem Testansatz eine Eichkurve zu erstellen.
- Nicht benötigte MP-Streifen sollten in der Hülle bei 2–8 °C trocken gelagert werden.

## 8.2. Testablauf

**Hinweis:** Es können andere als die empfohlenen Inkubationsbedingungen gewählt werden. Bei Abweichung vom vorliegenden Protokoll (z. B. Inkubationstemperatur 37 °C statt RT) ist der Anwender für die Validierung des Tests verantwortlich.

1. Je 100 µL der vorverdünnten (1:101) Patientenproben, gebrauchsfertigen Kalibratoren und Kontrollen in die entsprechenden Vertiefungen pipettieren.
2. Streifen bei RT für 30 Minuten inkubieren.
3. Platteninhalt verwerfen und Vertiefungen mit 3 x 300 µL gebrauchsfertigem Waschpuffer waschen. Rückstände von Waschpuffer sind durch Ausklopfen der Platten auf Fließpapier zu entfernen.
4. 100 µL Enzymkonjugat in alle Vertiefungen pipettieren.
5. Streifen bei RT für 30 Minuten inkubieren.
6. Waschen wie unter Punkt 3. beschrieben.
7. 100 µL TMB-Substrat pipettieren.
8. Streifen bei RT für 15 Minuten im Dunkeln inkubieren.
9. Reaktion durch Zugabe von 100 µL Stopplösung beenden, ca. 5 Minuten stehenlassen.
10. Inhalt der Vertiefungen kurz mischen und anschließend bei 450 nm messen. Als Blank wird gegen Luft gemessen. Es wird empfohlen, als Referenzwellenlänge 600–690 nm zu verwenden. Die Konzentrationen können graphisch anhand der Eichkurve oder mittels Computersimulation berechnet werden.

Die entwickelte Farblösung sollte innerhalb von 30 Minuten nach Zugabe der Stopplösung gemessen werden.

## 9. Auswertung und Interpretation

### Qualitativ

Zur qualitativen Testauswertung wird der Kalibrator 2 verwendet. Eine Quantifizierung nur unter Verwendung des Cut-off-definierenden Kalibrators ist nicht möglich.

### Quantitativ

Die Konzentration der Patientenproben wird mit gängigen Computerprogrammen berechnet. Für die Berechnung der Kalibrationskurve ist z. B. die 4-Parameter-Anpassung geeignet.

Optional: Die Extinktionswerte der 4 Kalibratoren und die entsprechenden Konzentrationen definieren die Punkte der Eichkurve. Für das Zeichnen der Eichkurve eignet sich besonders halblogarithmisches Millimeterpapier. Die Konzentration der Kalibratoren (s. Etikett bzw. QC-Zertifikat) wird auf der Abszisse (x-Achse) aufgetragen, die entsprechenden Extinktionswerte auf der Ordinate (y-Achse). Die Kalibrationskurve verläuft in der log-lin Auftragung sigmoid. Anhand dieser Kurve kann die Konzentration der Proben dann auf der Abszisse abgelesen werden.

### Grenzwerte

Anhand klinisch definierter Proben sowie gesunder Blutspender wurden nachstehende Normal- und Grenzwerte festgelegt:

negativ: < 25 IU/mL

positiv: > 25 IU/mL

### **Interpretation der Ergebnisse**

Kein Messergebnis darf für sich allein zu einer abschließenden Diagnose verwendet werden, sondern muss immer in Verbindung mit anderen Laborwerten und klinischen Befunden interpretiert werden.

Der Test erfasst spezifische Antikörper gegen doppelsträngige DNA. Hohe Testwerte sind häufig ein Hinweis auf ein aktives Stadium eines SLE. Allerdings sind isolierte, einzelne Messwerte weniger aussagekräftig als Messdaten, die anhand von Verlaufskontrollen gewonnen werden.

Die helicale Struktur der dsDNA auf der Festphase unterdrückt die Bindung von Anti-Purinbasen- bzw. Anti-Pyrimidinbasen-Autoantikörpern. Aufgrund der biochemischen Eigenschaften des DNA-Moleküls kann es in Einzelfällen zu Kreuzreaktionen mit anderen Autoantikörpern kommen, die gegen Phosphodiester- und/oder Phosphoriboseketten gerichtet sind.

### **Empfehlung**

Im Einzelfall sollten die Grenzwertbereiche an das jeweilige Patientenkollektiv angepasst werden.

## **10. Testcharakteristika**

Die Testcharakteristika des MASTAZYME™ dsDNA ELISAs wurden entsprechend den Vorgaben der IVD-Direktive der EU erstellt und bewegen sich im erwarteten Bereich. Auf Wunsch können diese Daten dem Anwender zur Verfügung gestellt werden.

## **11. Literatur**

1. The diagnostic value of antinuclear antibodies. Lukowsky A. et al., Hautarzt 46, 388-393, (1995).
2. Autoantibodies, Eds. J.B. Peter, Y. Shoenfeld, 1996, Elsevier
3. Pathogenic and diagnostic relevance of autoantibodies, Eds. K.Conrad, R.-L. Humbel, M. Meurer, Y.Shoenfeld, E.M.Tan, 1998, Pabst Science Publishers
4. Autoantikörper, K.Conrad, 1998, Pabst Science Publishers
5. Guidelines for clinical use of the antinuclear antibody test and tests for specific autoantibodies to nuclear antigens. American College of Pathologists; Kavanaugh A, Tomar R, Reveille J, Solomon DH, Homburger HA, Arch. Pathol. Lab. Med., 124 (1), 71-81 (2000).

<b>Contents</b>	<b>Page</b>
1. Intended use	10
2. Principle of the Test	10
3. Kit Contents	10
4. Materials Required but not Provided	11
5. Warnings and Precautions	11
6. Storage and Stability	12
7. Specimen Collection and Handling	12
8. Assay Procedure	12
9. Results and Interpretation	13
10. Assay Performance	14
11. References	14

## 1. Intended use

MASTAZYME™ dsDNA ELISA has been designed for the detection and quantification of specific antibodies against dsDNA respectively in serum and plasma.

This assay is intended for *in-vitro* diagnostic use only.

All laboratory test results should be interpreted in conjunction with other findings. Clinical judgement and additional tests should also be taken into account.

## 2. Principle of the Test

The ELISA test can be described in four stages:

### 2.1 Serum incubation

Specific antibodies bind to antigen on the solid phase to form a stable immune complex. After incubation for 30 minutes at room temperature the wells are washed with a prediluted wash buffer to remove all non-reactive serum components.

### 2.2 Conjugate incubation

The anti-human-IgG horseradish peroxidase conjugate is added to all wells. The conjugate binds to IgG antibodies on the solid phase antigen to form a stable sandwich. After incubating for 30 minutes at room temperature the excess conjugate is removed by washing all wells with wash buffer.

### 2.3 Substrate reaction and stopping

The TMB substrate is dispensed into each well and the peroxidase enzyme/substrate reaction forms a stable blue chromogen. The reaction and consequent colour development is stopped after a 15 minute incubation at room temperature by adding 0.5 N H<sub>2</sub>SO<sub>4</sub> to the wells. The shift in pH also causes the chromogen to change colour from blue to yellow.

### 2.4 Reading and interpretation

The colour intensity is read in a microtiter plate reader at 450 nm (recommended reference wavelength for bichromatic measurement: 600–690 nm). The intensity of the colour (OD) is directly proportional to the concentration of the specific antibody in the patient sample.

The results can be read from a calibration curve or with an electronic graphing package using a 4 parameter sigmoidal curve.

## 3. Kit Contents

The kits contains sufficient reagents for 12 x 8 = 96 determinations. Kit components should be stored at 2–8 °C. The expiry date is shown on the individual labels.

12	Microtiter strips	Single strips each with 8 break-apart wells coated with purified dsDNA respectively.
1 x	Frame holder	
4 x 1.5 mL	Calibrators 1–4	Stabilized human plasma containing antibodies against the above antigens and pre-diluted in buffer. Ready to use (RTU).

	Anti-dsDNA Concentration [IU/mL] as per WO/80
Calibrator 1	1
Calibrator 2 (Cut-off)	25
Calibrator 3	50
Calibrator 4	200
Positive Control	100
Negative Control	15

1 x 1.5 mL	Negative control	Stabilized human plasma, RTU, can be integrated into the calibration curve.
1 x 1.5 mL	Positive control	Stabilized human plasma, RTU, can be integrated into the calibration curve.
2 x 60 mL	Sample diluent	PBS solution with preservative, ready to use.
1 x 12 mL	Enzyme conjugate solution	HRP-labelled goat anti-human-IgG, ready to use.
1 x 12 mL	TMB substrate	3,3',5,5' Tetramethylbenzidine, ready to use.
1 x 12 mL	Stopping solution	0.5 N sulfuric acid, ready to use.
2 x 50 mL	Wash buffer 10 x concentrated	PBS/Tween buffer solution 10x concentrated. To be diluted 1:10 prior to use; the concentrate should be warmed up to 37 °C for 15 min to dissolve any crystals

#### 4. Materials Required but not Provided

- 5 µL-, 100 µL- and 500 µL micropipettes and a multichannel pipette (optional).
- Microtiter plate reader with a 450 nm filter (reference filter 600–690 nm)
- Microtiter plate washer (if washing manually: wash bottle)
- Reagent tubes for serum dilution
- Measuring cylinder
- Distilled water or water of similar or better quality

#### 5. Warning and Precautions

- For *in vitro* diagnostic use only! Do not ingest! Laboratory safety precautions should be followed. Do not eat, drink or smoke in the laboratory.
- All sera, plasma and buffers containing human biological material were found to be negative for Hepatitis B, C and HIV. Nevertheless, precautions such as the use of latex gloves must be taken.
- Serum and reagent spills should be cleaned with a disinfecting solution (e.g. sodium hypochlorite 5 %) and the refuse disposed of appropriately.
- All reagents should be brought to room temperature (18 to 24 °C) before starting the test procedure.
- Before pipetting, all reagents should be mixed thoroughly by gentle agitation. Vigorous shaking leading to formation of foam should be avoided.
- It is important to pipette with constant time intervals, so that all the wells of the microtiter plate have the same reaction conditions.
- When pipetting reagents out of the bottles care has to be taken that the stoppers are not contaminated. Furthermore, disposable pipette tips are strongly recommended to avoid cross-contamination. The contents of the bottles are usually sensitive to oxidation so should be opened only for a short time.
- To avoid cross contamination single use pipet tips should be used
- No reagents from different kit batches should be used and reagents should not be mixed with one another.

- All reagents should be used within the listed shelf life.
- In accordance with Good Laboratory Practice (GLP) or following ISO 9001 all laboratory equipment should be checked regularly regarding accuracy and precision. This refers e.g. to microliter pipets and washing or measurement (ELISA Reader) instrumentation.
- Certain reagents (especially the stop solution and substrate) are irritants- avoid contact with skin, eyes and mucosal membranes. In case of accident, rinse with water and seek medical attention. Clean all equipment after use to avoid secondary contact incidents.

## 6. Storage and Stability

Store all reagents at 2–8 °C.

The expiry date of each reagent is printed on the individual labels. Do not use any reagents after the expiry date has been exceeded.

The diluted wash buffer is stable for up to 4 weeks when stored at 2–8 °C. However ensure that this is at room temperature before testing.

Kits should be used within three months of opening.

## 7. Specimen Collection and Handling

Both serum and plasma (EDTA, heparin) can be used for testing. Samples can be stored at 2–8 °C for up to 3 days, although for longer storage periods they should be aliquoted and kept at -20 °C. Repeated freezing and thawing is contra-indicated. Lipaemic, haemolytic, icteric and bacterially contaminated samples can cause false positive or false negative results.

Patient sera should be prediluted 1:101 in sample diluent (e.g. 5 µL serum + 500 µL sample diluent) prior to testing.

## 8. Assay Procedure

### 8.1. Preparation of Reagents

Allow all kit components and specimens to reach room temperature (RT, 18–24 °C) prior to use and mix well.

**Washing buffer:** Dissolve any crystals which may be in the bottle by warming to 37 °C and then mix well.

Dilute the concentrated washing buffer 1:10 with distilled water (e.g. 50 mL buffer concentrate + 450 mL distilled water). Mix thoroughly.

- Strictly follow the instructions for reliable test performance. The user takes sole responsibility for any modifications to the test procedure.
- All reagents and samples must be brought to room temperature before use, but should not be left at this temperature for longer than necessary.
- A standard calibration curve should be established with each assay for quantitative results.
- Replace any unused microtiter strips in the resealable aluminium bag provided and store under dry conditions at 2–8 °C.

### 8.2. Assay Steps

Prepare a sufficient amount of microtiter wells for the calibrators, controls and samples.

**Note: Other incubation conditions are possible. However in case of modifications to the recommended test procedure (e.g. an incubation temperature of 37 °C instead of RT) assay performance should be validated by the user.**

1. Pipette 100 µL each of the diluted (1:101) samples and the ready to use calibrators and controls into the appropriate wells.

2. Incubate the test plate at room temperature for 30 minutes.
3. Discard the contents of the microwells and wash 3 times with 300 µL of diluted wash buffer. Afterwards remove residual wash buffer by gentle tapping of the microtiter plate onto a paper towel.
4. Pipette 100 µL of enzyme conjugate solution into each well.
5. Incubate the test plate for 30 minutes at room temperature.
6. Discard the contents of the microwells and wash 3 times with 300 µL of diluted wash buffer. Afterwards remove residual wash buffer by gentle tapping of the microtiter plate onto a paper towel.
7. Pipette 100 µL of TMB substrate into each well.
8. Incubate for 15 minutes in the dark (e.g. drawer) at room temperature.
9. Add 100 µL of stopping solution to each well.
10. After thorough mixing and wiping the bottom of the plate, read the optical density at 450 nm and calculate the results. Blank against air. A bichromatic measurement using a reference wavelength of 600–690 nm is recommended. Concentrations can be plotted with an electronic graphing package or by hand against the calibration curve.

The developed colour is stable for at least 30 minutes. Read optical densities during this time.

## 9. Results and Interpretation

### Qualitative

Calibrator 2 is used for qualitative interpretation of the assays, depending on specificity requirements to include or eliminate borderline results. A quantitative interpretation is not possible when only using calibrator 2 alone.

### Quantitative

The results can be calculated using a computer and suitable software program. In this case the results are calculated using a 4 parameter fit algorithm to produce a sigmoidal curve.

Optional: A standard curve is plotted by entering the mean absorbance of each calibrator on the Y-axis and the corresponding concentration on the X-axis using either log/log or linear/log graph paper. The concentration of the controls and specimen can then be read directly from the graph.

### Suggested normal values

In a normal range study using serum samples from healthy blood donors and disease-state sera the following normal and elevated ranges have been established. It is recommended that each laboratory should establish its individual normal ranges based on results obtained from the local population.

negative: < 25 IU/mL

positive: > 25 IU/mL

### Interpretation

Results can be affected by several patient factors. Clinical diagnosis and disease prognosis should not be based on a single test result. Other lab data and clinical observations must be taken into consideration before a conclusive diagnosis can be made.

The MASTAZYME™ dsDNA detects autoantibodies directed against double-stranded DNA. Strong positive samples usually indicate an acute phase SLE. Elevated values obtained from a single test are less indicative than results obtained from follow-up monitoring.

Binding of anti-purine and anti-pyrimidine base antibodies is suppressed through helical structure of solid phase dsDNA antigen. Biochemical structure of DNA molecules, however, may allow binding of antibodies directed against phosphodiester and/or phosphoribose structures present in the dsDNA antigen. Results can be affected by several patient factors. Clinical diagnosis and disease prognosis should not be based on a single test result. Other lab data and clinical observations must be taken into consideration before a conclusive diagnosis can be made.

### **Recommendation**

Each laboratory should establish its individual normal ranges based on results obtained from the local population.

## **10. Assay Performance**

The assay characteristics of the MASTAZYME™ dsDNA ELISAs have been established and assessed according the European IVD directive. Detailed validation data can be provided on special request.

## **11. References**

1. The diagnostic value of antinuclear antibodies. Lukowsky A. et al., Hautarzt 46, 388-393, (1995).
2. Autoantibodies, Eds. J.B. Peter, Y. Shoenfeld, 1996, Elsevier
3. Pathogenic and diagnostic relevance of autoantibodies, Eds. K.Conrad, R.-L. Humbel, M. Meurer, Y.Shoenfeld, E.M.Tan, 1998, Pabst Science Publishers
4. Autoantikörper, K.Conrad, 1998, Pabst Science Publishers
5. Guidelines for clinical use of the antinuclear antibody test and tests for specific autoantibodies to nuclear antigens. American College of Pathologists; Kavanaugh A, Tomar R, Reveille J, Solomon DH, Homburger HA, Arch. Pathol. Lab. Med., 124 (1), 71-81 (2000).

**Notizen / Notes / Note:**

**United Kingdom**

**Mast Group Ltd.  
MAST House Derby Road,  
Bootle, Mersey Side L20 1EA**

**Tel.: +44 151 933 7277  
Fax: +44 151 944 1332  
sales@mastgrp.com**



**Deutschland**

**Mast Diagnostica GmbH  
Feldstraße 20  
DÉ 23858 Reinfeld**

**Tel.: +49 4533 2007-0  
Fax: +49 4533 2007-68  
mast@mast-diagnostica.de**

**France**

**Mast Diagnostic  
115, Rue Jules Barni  
CS 91106  
80011 AMIENS CEDEX 1  
Tel.: + 33 3 22 80 80 67  
Fax: + 33 3 22 80 99 22  
info@mast-diagnostic.fr**

**[www.mastgrp.com](http://www.mastgrp.com)**