


■ CHROMagar[™] CCA

Plate Reading

- E.coli
- → metallic blue to violet
- Other coliforms
- → pink to red
- Other bacteria
- → colourless, inhibited



For the detection and enumeration of Escherichia coli and coliform bacteria in water according to the ISO 9308-1 norm

Background

Coliforms, Enterobacteriacae able to ferment lactose (lactose positive Enterobacteriacae), are bacteria present not only in human and warm blooded animals intestinal flora but also in the soil and water. Coliforms are proof of organic, environmental or faecal contamination. Faecal contamination, due to coliforms coming from animal waste, consists mainly of Escherichia coli and thermotolerant Klebsiella. Strict regulations exist for E. coli/coliform presence in water samples.

This can be explained by the importance of these germs in determining drinking water safety and process efficiency of treatment, storage and distribution.

Medium Performance

Simultanious detection and differentiation

CHROMagarTM CCA allows simultaneous detection and differentiation between E. coli and coliforms in one medium! This is helpful to determine if there is organic contamination (coliforms) or faecal contamination (E. coli). The use of this technique involves less work in comparison with traditional methods (MI Agar).

Easy to read

Easy counting of E. coli and coliform bacteria by the color.

3 Convenient

Especially suitable for waters with low bacterial numbers that will cause less than 100 total colonies, like drinking water, desinfected pool water, or finished water from drinking water treatment plants.

Medium Description

Powder Base	Total 21 F a/l
Powder base	Total
	Agar 14.9
	Peptone & Yeast extract
	NaCl
	Sodium dihydrogen phosphate
	Disodium hydrogen phosphate
	Sodium pyruvate
	Tryptophan1.0
	Sorbitol
	Tergitol® 15-S-7 0.15
	Chromogenic mix 0.5
	Storage at 15/30 °C - pH: 6.8 +/-0.2
	Shelf Life

Usual Samples	Water with low baterial background.
Procedure	Incubation 18-24h 36 °C +/- 2 °C
Ocionida Dublications on this grand out on the Company	

Scientific Publications on this product: available on www.CHROMagar.com
Please read carefully the instructions for use (IFU document) available on
www.CHROMagar.com

Manufacturer:

CHROMagar 4 place du 18 juin 1940 75006 Paris - Frankreich E-mail: CHROMagar@CHROMagar.com www.CHROMagar.com

Distributed by: Mast Diagnostica GmbH Feldstraße 20 DF-23858 Beinfeld

Tel.: +49 (0) 4533 2007 0 Fax: +49 (0) 4533 2007 68 E-Mail: mast@mast-diagnostica.de www.mast-group.com

Bestellinformationen

Produkt	Artikel-Nr.
CHROMagar™ CCA dry media, 500 g	15EF3435
CHROMagar™ CCA ready to use plates, 20 pcs.	201431