

Vitassay qPCR

Rotavirus

PCR en tiempo real para la detección cualitativa de Rotavirus en muestras humanas

Real-time PCR kit for the qualitative detection of Rotavirus in human samples



Uso previsto

Vitassay qPCR Rotavirus, permite la detección de rotavirus mediante PCR a tiempo real en muestras clínicas. Este producto está destinado para facilitar el diagnóstico diferencial de infecciones producidas por rotavirus.

Referencias

Vitassay qPCR Rotavirus 4x8 -well strip, low profile	7041017
Vitassay qPCR Rotavirus 4x8-well strip, high profile	7042017

Materiales/Reactivos suministrados

Código	Reactivo/Material	Color	Cantidad
7041S017/ 7042S017	Rotavirus strips low/high profile	-	4 tiras de 8 pocillos
7C017	Rotavirus Positive Control	rojo	1 vial
7001A	PCR grade water	blanco	1 vial x 1 mL
7002B	Resuspension buffer	verde	1 vial x 1,8 mL
7003N	Negative control	amarillo	1 vial x 1 mL
7004O	Tapas ópticas	-	4 tiras de 8 tapones

Condiciones de Transporte y conservación

- El transporte y almacenaje de los kits puede realizarse de 2-40°C hasta la fecha de caducidad indicada en la etiqueta.
- El control positivo resuspendido debe ser almacenado a -20°C. Para evitar ciclos repetidos de congelación y descongelación, se recomienda distribuir en alícuotas.
- Conservar los reactivos en oscuridad.

Material y equipamiento necesario, pero no proporcionado

- Kit de extracción de RNA
- Equipo de PCR a tiempo real (ver Adjunto I)
- Centrifuga para tubos de 1,5 mL
- Vórtex

- Micropipetas (1-20 µL, 20-200 µL)
- Puntas con filtro
- Guantes desechables sin polvo

Resumen

Los Rotavirus son virus pertenecientes a la familia *Reoviridae*, presentan una triple cápside proteica que rodea al genoma viral, el cual consiste en 11 segmentos de RNA de doble cadena que codifican seis proteínas estructurales virales (VP1-VP4, VP6 y VP7) y seis proteínas no estructurales (NSP1-NSP6). El genoma está contenido en el núcleo junto a tres proteínas VP1, VP2 y VP3. Rodeando este núcleo, hay una cápside interna compuesta por la proteína estructural principal VP6 que determina la clasificación antigenica del virus en siete grupos (A al G).

Aunque los serogrupos de Rotavirus A-C son patógenos para los humanos, los virus del grupo A (RVA) son los mayores causantes de gastroenteritis aguda (AGE) en niños menores de 5 años a nivel mundial. La infección está asociada con 500,000–600,000 muertes anuales, principalmente en países en vías de desarrollados. Rotavirus se transmite de persona a persona por vía oro-fecal debido al contacto con comida, agua u objetos contaminados. Los síntomas más característicos causados por estos virus son vómitos, diarrea y dolor abdominal.

Debido a que el gen *NSP3* contienen unas secuencias altamente conservadas, es la mejor diana para la detección de una gran variedad de genotipos de Rotavirus A.

Principio del test

Vitassay qPCR Rotavirus se basa en la amplificación a tiempo real de una región conservada específica del gen *NSP3* codificado por el genoma de rotavirus. El RNA viral extraído se transcribe a cDNA utilizando un *primer* específico mediante un paso de transcripción inversa seguido, de la reacción en cadena de la polimerasa. La presencia de virus rotavirus se detecta mediante un aumento de fluorescencia durante la reacción, tras la hidrólisis de la sonda fluorescente.

El ensayo está basado en la actividad 5' nucleasa que utiliza dos *primers* y una sonda de hidrólisis fluorogénica para detectar la acumulación de la secuencia diana amplificada durante la reacción de PCR. Cuando la polimerasa comienza a extender los primers, la sonda es hidrolizada mediante su actividad exonucleasa 5'- 3' produciendo la separación espacial del fluoróforo y el *quencher*. El aumento de la señal fluorescente resultante es proporcional a la cantidad de producto amplificado en la muestra y es detectado mediante un equipo de PCR en tiempo real.

Vitassay qPCR Rotavirus se trata de un ensayo listo para usar que contiene en cada pocillo todos los reactivos necesarios en formato estabilizado para llevar a cabo la PCR

en tiempo real. Además, un control interno permite la detección de una posible reacción de inhibición. La amplificación de la secuencia diana es detectada en el canal FAM mientras que el control interno (CI) se detecta en el canal HEX, VIC o JOE (según el equipo utilizado).

Precauciones

- Diseñado para uso profesional de diagnóstico *in vitro*.
- No utilizar el kit después de la fecha de caducidad.
- No mezclar reactivos de otros kits y/o diferentes lotes.
- No utilizar si el kit tiene signos de haber sido abierto o manipulado.
- No utilizar el kit si el material desecante de los diferentes sobres de aluminio está dañado o no está.
- Se recomienda proteger los tubos de la humedad ya que una exposición prolongada puede afectar al rendimiento del producto.
- Trabajar siguiendo las Buenas Prácticas de Laboratorio. Use ropa protectora, guantes desechables, gafas y mascarilla.
- No comer, beber o fumar en la zona de trabajo.
- Es importante seguir un flujo de trabajo en el laboratorio unidireccional: Área de Extracción, Área de Amplificación y Detección. No retornar muestras, equipos ni reactivos a un área anterior.
- Las muestras y todo material en contacto con ellas se deben tratar como potencialmente infecciosos y se deben gestionar según la legislación sobre residuos sanitarios nacional. Tome las precauciones necesarias durante la recogida, almacenamiento, tratamiento y eliminación de muestras.
- Se recomienda la descontaminación periódica de los equipos usados habitualmente, especialmente micropipetas, y de las superficies de trabajo.

Procedimiento

Toma de muestra, preparación y extracción de RNA

Realizar el pretratamiento y el aislamiento de los ácidos nucleicos utilizando un sistema manual o automático compatible con ensayos de PCR en tiempo real. Seguir las instrucciones de uso del fabricante. Los siguientes kits de extracción han sido validados:

QIAamp MinElute Virus Spin Kit (QIAGEN)

QIAamp Viral RNA Mini Kit (QIAGEN)

Invisorb® Spin Universal Kit (Stratec).

NucleoSpin® RNA Virus (Machery Nagel).

RIDA® Xtract (R-biopharm).

Maxwell® RSC Blood DNA Kit, utilizando Maxwell® 16 instrument (Promega).

NucliSENS® EasyMAGTM platform (bioMérieux).

Preparación del control positivo

Reconstituir el contenido liofilizado del Rotavirus Positive Control (tubo rojo) con 100 µL de agua ultrapura (PCR grade water, tubo blanco). Mezclar hasta conseguir una suspensión homogénea con ayuda del vórtex. Después del primer uso, dispensar en alícuotas para evitar repetidos ciclos de congelación-descongelación y almacenarlo a -20°C.

El control positivo contiene una gran cantidad de copias molde y el riesgo de contaminación es elevado. Por lo tanto, se recomienda abrir y manipular en un área del laboratorio separada de los otros componentes del kit y de las muestras a analizar.

Preparación de la reacción

- Preparar el número de pocillos necesarios incluyendo muestras y controles (un control positivo y uno negativo).
- Retirar el sello de aluminio que protege las tiras.
- Pipetear 15 µL de la solución de resuspensión (tubo verde) y añadirlos en cada pocillo.
- Pipetear 5 µL de DNA extraído, Control Negativo (tubo amarillo) y Control positivo (tubo rojo) y añadirlos en los pocillos correspondientes.
- Cerrar los pocillos con los tapones suministrados. Centrifugar brevemente (opcional).
- Colocar las tiras en el equipo de PCR a tiempo real.

Programación del termociclador

Configurar el termociclador siguiendo las siguientes instrucciones:

Etapa	Temperatura	Tiempo	Ciclos
Retrotranscripción	45°C	15 min	1
Desnaturalización inicial	95°C	2 min	1
Desnaturalización	95°C	10 seg	
Hibridación/Elongación (Recogida de datos*)	60°C	50 seg	45

Los datos de fluorescencia deben recogerse durante la etapa de hibridación (*) a través de los canales FAM (rotavirus) y los canales HEX, JOE o VIC (Control Interno). En los termocicladores Applied Biosystems 7500 Fast, y Stratagene Mx3005P™ comprobar que la opción del control pasivo ROX esta desactivada (ver Adjunto II).

Análisis e interpretación de resultados

El análisis de las muestras se realiza con el software propio del equipo de PCR a tiempo real de acuerdo con las instrucciones de uso del fabricante.

Antes de analizar el resultado de las muestras clínicas debe validarse el resultado de los controles:

Control positivo

El control positivo utilizado en cada serie debe mostrar una curva de amplificación en los canales de rotavirus (FAM).

Control negativo

El control negativo incluido en cada serie debe mostrar la ausencia de señal de FAM.

El experimento es inválido si hay señal de amplificación en el control negativo o ausencia de la señal en el control positivo. El ensayo se debe de repetir.

Con la ayuda de la siguiente tabla, analizar los resultados:

Rotavirus (FAM)	Control Interno	Control Negativo	Control Positivo	Interpretación
+	+/-	-	+	Rotavirus Positivo
-	+	-	+	Rotavirus Negativo
+	+	+	+	Inválido
-	-	-	-	Inválido

Positivo (+): Señal de amplificación

Negativo (-): No hay señal de amplificación

Si las muestras negativas para todas las bacterias no muestran un resultado positivo para el control interno, se debe repetir el ensayo diluyendo la muestra original 1:10 o repetir la extracción de los ácidos nucleicos debido a posibles problemas causados por inhibidores de PCR.

Control de Calidad

Con el fin de confirmar el correcto funcionamiento de la técnica de diagnóstico molecular, se incluye un Control Interno (CI) en cada reacción. Además, un control positivo y uno negativo se debe de incluir en cada ensayo para una correcta interpretación de los resultados.

Características técnicas

Sensibilidad y especificidad clínica

Un total de 135 muestras fecales provenientes de pacientes sintomáticos fueron evaluadas con los siguientes test de diagnóstico molecular: Vitassay qPCR Rotavirus, RIDA®GENE Rotavirus/Adenovirus Duplex, RIDA®GENE Viral Stool Panel II (r-Biopharm) y Rotavirus/Norovirus/Astrovirus Real-TM (Sacace). Rotavirus fue detectado en 34 muestras mediante el test Vitassay qPCR Rotavirus. De estas muestras, solamente una no pudo ser detectada por los otros tests.

Además, Vitassay qPCR Rotavirus fue evaluado con el programa externo QCMD 2017 viral gastroenteritis EQA pilot study (GastroV17). Este programa consistía en un panel de 9 muestras en SFM positivas o negativas para el patógeno de estudio. Vitassay qPCR Rotavirus detectó correctamente todas las muestras presentes en el panel.

El conjunto de estos resultados nos indica la alta sensibilidad y especificidad para detectar Rotavirus utilizando el kit de diagnóstico molecular Vitassay qPCR Rotavirus.

Sensibilidad analítica

La sensibilidad analítica fue determinada a partir de diluciones seriadas (1:10) del RNA molde de las diferentes bacterias (10^7 - 10^1 copias/reacción). Este ensayo tiene un límite de detección de ≥ 10 copias de RNA por reacción.

Especificidad analítica

La especificidad analítica para la detección de rotavirus fue confirmada probando un panel compuesto por diferentes microorganismos.

No se observaron reacciones cruzadas de rotavirus entre ninguna de las especies:

Cepas empleadas en las pruebas de reactividad cruzada		
Adenovirus 40/41	<i>Cryptosporidium parvum</i>	<i>Salmonella paratyphi A</i>
Astrovirus Genotype I-VIII	<i>Entamoeba histolytica</i>	<i>Salmonella paratyphi B</i>
Norovirus GI and GII	<i>Enterococcus faecalis</i>	<i>Salmonella typhimurium</i>
<i>Aeromonas hydrophila</i>	<i>Enterotoxigenic E. coli (ETEC)</i>	<i>Salmonella bongori</i>
<i>Arcobacter butzleri</i>	<i>Enteropathogenic E. coli (EPEC)</i>	<i>Salmonella enteritidis</i>
<i>Bacteroides fragilis</i>	<i>Giardia intestinalis</i>	<i>Salmonella enterica</i>
<i>Campylobacter lari</i>	<i>Helicobacter pylori</i>	<i>Salmonella pullorum</i>
<i>Campylobacter fetus</i>	<i>Helicobacter hepaticus</i>	<i>Salmonella gallinarum</i>
<i>Campylobacter coli</i>	<i>Helicobacter cinaedi</i>	<i>Serratia liquefaciens</i>
<i>Campylobacter jejuni</i>	<i>Helicobacter heilmannii</i>	<i>Shigella flexneri</i>
<i>Campylobacter upsaliensis</i>	<i>Klebsiella oxytoca</i>	<i>Shigella dysenteriae</i>
<i>Candida albicans</i>	<i>Listeria monocytogenes</i>	<i>Staphylococcus aureus</i>
<i>Citrobacter freundii</i>	<i>Pseudomonas aeruginosa</i>	<i>Vibrio parahaemolyticus</i>
<i>Clostridium difficile</i>	<i>Proteus vulgaris</i>	<i>Y. enterocolitica O:3 / O:9</i>
<i>Clostridium perfringens</i>	<i>Salmonella typhi</i>	

Reactividad analítica

Vitassay qPCR Rotavirus para Rotavirus ha sido evaluado frente a rotavirus A humano cepa ATCC ® VR-2018TM y los genotipos Rotavirus G1P[8], G2P[4], G2P[8], G3P[8], G4P[8], G1+G3P[8], G6P[8], G9P[8], G10+G12P[8], G12P[8] y cepa G-UDP[8], obteniéndose un resultado positivo para todos ellas.

Termocicladores compatibles

Vitassay qPCR Rotavirus, ha sido probado en los siguientes equipos:

- Cobas Z480 (Roche)
- 7500 Fast Real-Time PCR System (Applied Biosystems) ^{II}
- StepOne™ Real-Time PCR System (Applied Biosystems) ^{II}
- CFX96 ™ Real-Time PCR Detection System (Bio-Rad)
- AriaMx Real-Time PCR System (Agilent Technologies)
- DTlite Real-Time PCR System (DNA-Technology)
- DTPrime Real Time Detection Thermal Cycler (DNA-Technology)
- Rotor-Gene® Q (Qiagen)^I
- SmartCycler® (Cepheid)^I

I: Para los equipos Rotor-Gene® Q y SmartCycler® el producto debe ser reconstituido y trasvasado a los tubos específicos de cada uno de los equipos.

II: En el caso de utilizar el equipo Applied Biosystems 7500 Fast con tiras, se recomienda colocar el soporte adecuado para los tubos (Ref. PN 4388506).

Limitaciones

- Este test proporciona un diagnóstico preliminar de infección por rotavirus. Todos los resultados obtenidos deben ser interpretados por un especialista junto con la información clínica y los hallazgos de laboratorio disponibles.
- Este ensayo ha sido probado en muestras fecales humanas. El uso de otras muestras no se ha establecido.
- El correcto funcionamiento de la prueba depende de la calidad de la muestra; el RNA debe ser extraído de forma adecuada de las muestras clínicas humanas. Una forma inadecuada de recolección, almacenaje y/o transporte de las muestras puede dar lugar a falsos negativos.
- Se puede detectar un bajo número de copias del RNA molde diana por debajo del límite de detección, pero los resultados pueden no ser reproducibles.
- Existe la posibilidad de falsos positivos debido a la contaminación cruzada con las diferentes bacterias, ya sea por muestras que contienen altas concentraciones de RNA molde diana o por contaminación por arrastre a partir de productos de PCR de reacciones anteriores.

Adjunto I: Compatibilidad de los termocicladores a tiempo real más usuales

Los termocicladores más usuales se enumeran en la siguiente tabla separados por tipo de tubo. Si no encuentra su termociclador, póngase en contacto con su proveedor:

Termocicladores con bloque de bajo perfil	Termocicladores con bloque de alto perfil
Agilent Technologies	Abbott
AriaMx Real-Time PCR System	Abbott m2000 RealTime System
Applied Biosystems	Applied Biosystems
7500 Fast Real-Time PCR System	7300 Real-Time PCR System
7500 Fast Dx Real-Time PCR System	7500 Real-Time PCR System
QuantStudio™ 12K Flex 96-well Fast	7900 HT Real-Time PCR System
QuantStudio™ 6 Flex 96-well Fast	ABI PRISM 7000
QuantStudio™ 7 Flex 96-well Fast	ABI PRISM 7700
QuantStudio™ 5 Real-Time PCR System	QuantStudio™ 12K Flex 96-well
QuantStudio™ 3 Real-Time PCR System	QuantStudio™ 6 Flex 96-well
StepOne Plus™ Real-Time PCR System	QuantStudio™ 7 Flex 96-well
StepOne™ Real-Time PCR System	QuantStudio™ 5 Real-Time PCR
ViiA™ 7 Fast Real-Time PCR System	QuantStudio™ 3 Real-Time PCR System
BIONEER	ViiA™ 7 Real-Time PCR
Exicycler™ 96	Analytik Jena Biometra
Bio-Rad	TOptical
CFX96™ Real-Time PCR Detection System	qTOWER 2.0
Mini Opticon™ Real-Time PCR Detection System	BIONEER
Cepheid	Exicycler™ 96
SmartCycler®	Bio-Rad
Qiagen	CFX96™ Deep Well Real-Time PCR Detection
Rotor-Gene® Q	iCycler iQ™ Real-Time PCR
Roche	iCycler iQ™ 5 Real-Time PCR
LightCycler® 480 Real-Time PCR System	MyIQ™ Real-Time PCR Detection System
LightCycler® 96 Real-Time PCR System	MyIQ™ 2Real-Time PCR Detection System
Cobas z480 Analyzer	Cepheid
	SmartCycler®
	DNA-Technology
	DTlite Real-Time PCR System*
	DTprime Real-time Detection Thermal Cycler*
	Eppendorf
	Masterycler™ ep realplex
	Qiagen
	Rotor-Gene® Q
	Stratagene / Agilent Technologies
	Mx3000P™ Real Time PCR System
	Mx3005P™ Real Time PCR System

* Ver Adjunto III para configurar los valores de exposición

Adjunto II: Canales de detección de los equipos a tiempo real

Los canales de fluorescencia de algunos de los termocicladores a tiempo real más comunes se especifican en la siguiente tabla:

Termociclador	Canal Vitassay	Canal de Detección	Observaciones
Bio-Rad CFX96™	FAM	FAM	Opción del control pasivo ROX desactivada
	HEX	HEX	
	ROX	ROX	
	Cy5	Cy5	
ABI 7500 Applied Biosystems	FAM	FAM	Opción del control pasivo ROX desactivada
	HEX	VIC	
	ROX	ROX	
	Cy5	Cy5	
Roche Lightcycler®480II	FAM	465/510	Se requiere compensación de color
	HEX	533/580	
	ROX	533/610	
	Cy5	618/660	
Smartcycler® Cepheid	FAM	Channel 1	
	HEX	Channel 2	
	ROX	Channel 3	
	Cy5	Channel 4	
Abbott m2000rt	FAM	FAM	
	HEX	VIC	
	ROX	ROX	
	Cy5	Cy5	
Mx3000P™ Mx 3005P™ Stratagene	FAM	FAM	Opción del control pasivo ROX desactivada
	HEX	VIC	
	ROX	ROX	
	Cy5	Cy5	
AriaMx Agilent	FAM	FAM	
	HEX	HEX	
	ROX	ROX	
	Cy5	Cy5	
Rotor-Gene®Q Qiagen	FAM	Green	Al configurar los canales, presione el botón "Gain Optimisation" y después vaya a "Optimise Acquiring". La fluorescencia del apartado Target Sample Range tiene que estar entre 5 y 10 FI para cada canal. Además, marque la opción "Perform Optimisation Before 1st Acquisition".
	HEX	Yellow	
	ROX	Orange	
	Cy5	Red	
Exicycler™ 96 BIONEER	FAM	FAM	
	HEX	JOE	
	ROX	ROX	
	Cy5	Cy5	

Adjunto III: Configuración valores de exposición

Los valores de exposición de algunos termocicladores se deben ajustar para su correcto funcionamiento. Establecer los valores de exposición de la siguiente manera:

Termociclador	Canal Vitassay	Valor de exposición
DTlite Real-Time PCR System (DNA-Technology)	FAM	500
	HEX	500
	ROX	500
	Cy5	500
DTprime Real-time Detection Thermal Cycler (DNA-Technology)	FAM	500
	HEX	1000
	ROX	1000
	Cy5	1000

-
-

Intended use

Vitassay qPCR Rotavirus allow the detection and differentiation of Rotavirus by real-time PCR in clinical samples. The product is intended for use in the diagnosis of Rotavirus infections alongside clinical data of the patient and other laboratory tests outcomes.

References

Vitassay qPCR Rotavirus 4x8 -well strip, low profile	7041017
Vitassay qPCR Rotavirus 4x8-well strip, high profile	7042017

Materials/reagents provided

Reference	Reagent/Material	Colour	Amount
7041S017/ 7042S017	Rotavirus strips low/high profile	-	4x8-well strip
7C017	Rotavirus Positive Control	red	1 vial
7001A	PCR grade water	white	1 vial x 1 mL
7002B	Resuspension Buffer	green	1 vial x 1,8 mL
7003N	Negative control	yellow	1 vial x 1 mL
7004O	Optical caps	-	4x8 cap strip

Transport and storage

- The reagents and the test can be shipped and stored at 2-40°C until expiration date stated in the label.
- The resuspended positive control should be stored at -20°C. In order to avoid repeated freeze/thaw cycles, it is recommended to distribute the content in different aliquots.
- Keep all reagents in the darkness.

Additional equipment and material required

- RNA extraction kit
- Real-time PCR instrument (thermocycler) (Attached I)
- Centrifuge for 1.5 mL tubes
- Vortexer
- Micropipettes (1-20 µL, 20-200 µL)
- Filter tips
- Powder-free disposal gloves

Summary

Rotaviruses are triple-layered viruses that belong to the *Reoviridae* family and contain a genome of 11 segments of double-stranded (ds) RNA that encode six structural viral proteins (VP1–VP4, VP6, and VP7) and six nonstructural proteins (NSP1–NSP6). The genome is contained in the core with three proteins VP1, VP2 and VP3. Surrounding this core, there is an inner capsid composed for the main structural protein VP6 which determines the antigen classification of virus in seven groups (A to G).

Although rotavirus serogroups A-C are human pathogens, Group A rotavirus (RVA) infection is the leading cause of acute gastroenteritis (AGE) in children aged <5 y worldwide and is associated with 500,000–600,000 deaths annually, predominantly in developing countries. Rotavirus is transmitted from person-to-person by fecal-oral route due to food, objects or water contaminated. The typical symptoms caused by these viruses are vomiting, watery diarrhea and abdominal pain.

Due to the highly conserved sequences of the *NSP3* gene, this gene has been shown to be the best target for detection of a wide variety of RVA genotypes.

Principle of the test

Vitassay qPCR Rotavirus is designed for the identification of rotavirus virus in clinical specimens to aid in the assessment of infections caused by this virus.

Vitassay qPCR Rotavirus is based on the real-time amplification of specific conserved fragments of the *NSP3* gene encoded by the Rotavirus virus genome. The viral RNA extracted is transcribed into cDNA using a specific primer by reverse transcription step followed immediately in the same well by polymerase chain reaction. The presence of Rotavirus virus is detected by an increase in observed fluorescence during the reaction upon hydrolysis of the fluorescent probe.

Vitassay qPCR Rotavirus test is a ready-to-use test which contains in each well all the necessary reagents for real-time PCR assay in a stabilized format. In addition, an internal control allows the detection of a possible reaction inhibition. The amplification of the Rotavirus target sequence is detected through the FAM channel and whereas the internal control (IC) in HEX, VIC or JOE channel (depending on the equipment used select the proper detection channel, see Attached II).

Precautions

- For *in vitro* diagnostic use.
- Do not use after expiration date.
- Do not mix components from different kits and/or lots.
- Do not use if package is open or damaged.

- Do not use the kit if the desiccant from the different reagents is not present or broken or if the foil has been broken or damaged.
- Protect the kit against humidity. Prolonged exposure to humidity may affect product performance.
- Follow Good Laboratory Practices. Wear protective clothing, use disposal gloves, goggles and mask.
- Do not eat, drink or smoke in the working area.
- The laboratory process must be one-directional, it should begin in the Extraction Area and then move to the Amplification and Detection Areas. Do not return samples, equipment and reagents to the area in which the previous step was performed.
- Specimens and reagents/materials that have been exposed to them must be treated as potentially infectious. Take necessary precautions during the collection, storage, treatment and disposal of samples.
- Regular decontamination of commonly used equipment is recommended, especially micropipettes and work surfaces.

Procedures

Specimen collection, processing and RNA extraction

For pre-treatment and nucleic acid isolation, it is recommended to use your optimized manual or automatic system compatible with real-time PCR technology. The assay has been validated with the following extraction kits:

QIAamp MinElute Virus Spin Kit (QIAGEN)

QIAamp Viral RNA Mini Kit (QIAGEN)

Invisorb® Spin Universal Kit (Stratec).

NucleoSpin® RNA Virus (Machery Nagel).

RIDA® Xtract (R-biopharm).

Maxwell® RSC Blood DNA Kit, using the Maxwell® 16 instrument (Promega).

NucliSENS® EasyMAGTM platform (bioMérieux).

Positive control preparation

Reconstitute the lyophilized Rotavirus Positive Control (red tube) with 100 µL of PCR grade water (white tube). To ensure a complete resuspension, vortex the tube thoroughly. After first use, dispense into aliquots in order to avoid multiple freeze-thaw cycles, and store them at -20°C.

This component contains high copies number template and is a very significant contamination risk. Therefore, we recommend open and manipulate it in a separate laboratory area away from the other components and samples.

Reaction setup

- Separate the number of required reactions including samples and controls. Remember that one positive and one negative control must be included in each run. Peel off protective aluminium seal from the strips.
- Pipette 15 µL of Resuspension buffer (green tube) into each well.
- Pipette 5 µL of RNA sample, negative and positive controls into each well.
- Cover the wells with the caps provided. Spin down briefly (optional).
- Place the strips in the Real-time PCR instrument.

Programme your thermocycler

Set your thermocycler following the conditions below:

Step	Temperature	Time	Cycles
Reverse transcription	45°C	15 min	1
Initial denaturation	95°C	2 min	1
Denaturation	95°C	10 sec	
Annealing/Extension (Data collection*)	60°C	50 sec	45

Set the fluorescence data collection during the extension step (*) through the FAM (rotavirus) and HEX, JOE or VIC channels (Internal Control (IC)). If you use the Applied Biosystems 7500 Fast or the Stratagene Mx3005P™ check that passive reference option ROX is none. (Attached II)

Analysis and interpretation of results

The analysis of the results is done by the software itself of the used real-time PCR system following manufacturer's instructions.

For a valid diagnostic test run, the following control conditions must be met:

Positive control

The positive controls used in each run, must show an amplification curve for FAM (Rotavirus) which validates the reaction.

Negative control

The negative controls included in each run, must show the absence of signal in FAM which validates the reaction.

The experiment seems to be failed if there is signal of amplification in negative control or absence of signal in the positive well. The assay should be repeated.

The result interpretation is summarized in the following table:

Rotavirus (FAM)	Internal Control	Negative Control	Positive Control	Interpretation
+	+/-	-	+	Rotavirus Positive
-	+	-	+	Rotavirus Negative
+	+	+	+	Experiment fail
-	-	-	-	Experiment fail

(+) Positive: Amplification signal

(-) Negative: No amplification signal

If the negative samples not show a positive result for the internal control, they should be retested from the diluted original sample 1:10 or the nucleic acid extraction has to be repeated due to possible problems caused by PCR inhibitors.

Quality Control

In order to confirm the appropriate performance of the molecular diagnostic technique, an Internal Control (IC) is included in each reaction. Besides, a positive and a negative control must be included in each assay to interpret the results correctly.

Performance evaluation

Clinical sensitivity and specificity

Overall, 135 fecal samples from symptomatic patients were tested using different molecular diagnostic test: Vitassay qPCR Rotavirus; RIDA®GENE Rotavirus/Adenovirus Duplex; RIDA®GENE Viral Stool Panel II (r-Biopharm) and Rotavirus/Norovirus/Astrovirus Real-TM (Sacace). Rotavirus was detected in 34 samples by Vitassay qPCR Rotavirus. Among these, only one sample was not able to be detected by the other tests.

In addition, Vitassay qPCR Rotavirus was evaluated with the external programme QCMD 2017 viral gastroenteritis EQA pilot study (GastroV17). This programme consisted of a panel of 9 samples in SFM positive or negative for the studied pathogen. Vitassay qPCR Rotavirus correctly detected all the samples present in the panel.

These results indicate the high sensitivity and specificity to detect Rotavirus using the molecular diagnostic kit Vitassay qPCR Rotavirus.

Analytical sensitivity

The analytical sensitivity was determined by analysis of 10-fold dilution series of Rotavirus templates ranging from 10^7 to 10^1 copies/rxn. This assay has a detection limit of ≥ 10 RNA copies per reaction.

Analytical specificity

The analytical specificity for Rotavirus was tested within the panel of different microorganisms.

No cross-reactivity of Rotavirus was seen between any of the species:

Cross-reactivity assay		
Adenovirus 40/41	<i>Cryptosporidium parvum</i>	<i>Salmonella paratyphi A</i>
Astrovirus Genotype I-VIII	<i>Entamoeba histolytica</i>	<i>Salmonella paratyphi B</i>
Norovirus GI and GII	<i>Enterococcus faecalis</i>	<i>Salmonella typhimurium</i>
<i>Aeromonas hydrophila</i>	<i>Enterotoxigenic E. coli (ETEC)</i>	<i>Salmonella bongori</i>
<i>Arcobacter butzleri</i>	<i>Enteropathogenic E. coli (EPEC)</i>	<i>Salmonella enteritidis</i>
<i>Bacteroides fragilis</i>	<i>Giardia intestinalis</i>	<i>Salmonella enterica</i>
<i>Campylobacter lari</i>	<i>Helicobacter pylori</i>	<i>Salmonella pullorum</i>
<i>Campylobacter fetus</i>	<i>Helicobacter hepaticus</i>	<i>Salmonella gallinarum</i>
<i>Campylobacter coli</i>	<i>Helicobacter cinaedi</i>	<i>Serratia liquefaciens</i>
<i>Campylobacter jejuni</i>	<i>Helicobacter heilmannii</i>	<i>Shigella flexneri</i>
<i>Campylobacter upsaliensis</i>	<i>Klebsiella oxytoca</i>	<i>Shigella dysenteriae</i>
<i>Candida albicans</i>	<i>Listeria monocytogenes</i>	<i>Staphylococcus aureus</i>
<i>Citrobacter freundii</i>	<i>Pseudomonas aeruginosa</i>	<i>Vibrio parahaemolyticus</i>
<i>Clostridium difficile</i>	<i>Proteus vulgaris</i>	<i>Y. enterocolitica O:3 / O:9</i>
<i>Clostridium perfringens</i>	<i>Salmonella typhi</i>	

Analytical reactivity

Analytical reactivity of Vitassay qPCR Rotavirus for Rotavirus was evaluated against human Rotavirus A strain ATCC ® VR-2018TM and Rotavirus genotypes G1P[8], G2P[4], G2P[8], G3P[8], G4P[8], G1+G3P[8],]G6P[8], G9P[8], G10+G12P[8], G12P[8] and G-UDP[8] strains, showing positive result.

Compatibles real-time PCR equipment

Vitassay qPCR Rotavirus has been validated on the following equipments:

- Cobas Z480 (Roche)
- 7500 Fast Real-Time PCR System (Applied Biosystems) ^{II}
- StepOne™ Real-Time PCR System (Applied Biosystems) ^{II}
- CFX96™ Real-Time PCR Detection System (Bio-Rad)
- AriaMx Real-Time PCR System (Agilent Technologies)
- DTlite Real-Time PCR System (DNA-Technology)
- DTPrime Real Time Detection Thermal Cycler (DNA-Technology)
- Rotor-Gene® Q (Qiagen)^I
- SmartCycler® (Cepheid)^I

I: For Rotor-Gene® Q and SmartCycler® thermocyclers the product should be reconstituted following the procedure and transferred into specific Rotor-Gene® Q and/or SmartCycler tubes.

II: When using the Applied Biosystems 7500 Fast with strips it is recommend placing a plate holder for the tubes (Ref. PN 4388506).

Limitations

- This test provides a presumptive diagnosis of rotavirus infection. All results must be interpreted together with other clinical information and laboratory findings available to the physician.
- This assay was tried with human faecal samples. The use of other samples has not been established.
- The quality of the test depends on the quality of the sample; proper RNA from clinical specimens must be extracted. Unsuitable collection, storage and/or transport of specimens may give false negative results.
- Extremely low levels of target below the limit of detection may be detected, but results may not be reproducible.
- There is a possibility of false positive results due to cross-contamination by the different bacteria, either samples containing high concentrations of target RNA or contamination due to PCR products from previous reactions.

Attached I: Compatibility of the most common real-time PCR equipment

The most common thermocyclers are listed in the following table separated by tube type. If you do not find your thermocycler, please contact with your supplier.

Low profile Block Thermocyclers	High profile Block Thermocyclers
Agilent Technologies	Abbott
AriaMx Real-Time PCR System	Abbott m2000 RealTime System
Applied Biosystems	Applied Biosystems
7500 Fast Real-Time PCR System	7300 Real-Time PCR System
7500 Fast Dx Real-Time PCR System	7500 Real-Time PCR System
QuantStudio™ 12K Flex 96-well Fast	7900 HT Real-Time PCR System
QuantStudio™ 6 Flex 96-well Fast	ABI PRISM 7000
QuantStudio™ 7 Flex 96-well Fast	ABI PRISM 7700
QuantStudio™ 5 Real-Time PCR System	QuantStudio™ 12K Flex 96-well
QuantStudio™ 3 Real-Time PCR System	QuantStudio™ 6 Flex 96-well
StepOne Plus™ Real-Time PCR System	QuantStudio™ 7 Flex 96-well
StepOne™ Real-Time PCR System	QuantStudio™ 5 Real-Time PCR
ViiA™ 7 Fast Real-Time PCR System	QuantStudio™ 3 Real-Time PCR System
BIONEER	ViiA™ 7 Real-Time PCR
Exicycler™ 96	Analytik Jena Biometra
Bio-Rad	TOptical
CFX96™ Real-Time PCR Detection System	qTOWER 2.0
Mini Opticon™ Real-Time PCR Detection System	BIONEER
Cepheid	Exicycler™ 96
SmartCycler®	Bio-Rad
Qiagen	CFX96™ Deep Well Real-Time PCR Detection
Rotor-Gene® Q	iCycler iQ™ Real-Time PCR
Roche	iCycler iQ™ 5 Real-Time PCR
LightCycler ®480 Real-Time PCR System	MyiQ™ Real-Time PCR Detection System
LightCycler ®96 Real-Time PCR System	MyiQ™ 2Real-Time PCR Detection System
Cobas z480 Analyzer	Cepheid
	SmartCycler®
	DNA-Technology
	DTlite Real-Time PCR System
	DTprime Real-time Detection Thermal Cycler
	Eppendorf
	Masterycler™ ep realplex
	Qiagen
	Rotor-Gene® Q
	Stratagene / Agilent Technologies
	Mx3000P™ Real Time PCR System
	Mx3005P™ Real Time PCR System

* See Attached III to configure exposure settings.

Attached II: Detection channels of most common real time PCR equipment

The fluorescence detection channels of some of most common Real Time PCR Thermocyclers are specified in the following table:

REAL-TIME PCR THERMOCYCLER	VITASSAY CHANNEL	DETECTION CHANNEL	OBSERVATIONS
Bio-Rad CFX96™	FAM	FAM	
	HEX	HEX	
	ROX	ROX	
	Cy5	Cy5	
ABI 7500 Applied Biosystems	FAM	FAM	Passive reference option ROX is none
	HEX	VIC	
	ROX	ROX	
	Cy5	Cy5	
Roche Lightcycler®480II	FAM	465/510	Colour Compensation required
	HEX	533/580	
	ROX	533/610	
	Cy5	618/660	
Smartcycler® Cepheid	FAM	Channel 1	
	HEX	Channel 2	
	ROX	Channel 3	
	Cy5	Channel 4	
Abbott m2000rt	FAM	FAM	
	HEX	VIC	
	ROX	ROX	
	Cy5	Cy5	
Mx3000P™ Mx 3005P™ Stratagene	FAM	FAM	Passive reference option ROX is none
	HEX	VIC	
	ROX	ROX	
	Cy5	Cy5	
AriaMx Agilent	FAM	FAM	
	HEX	HEX	
	ROX	ROX	
	Cy5	Cy5	
Rotor-Gene®Q Qiagen	FAM	Green	In the Channel Setup, click on the "Gain Optimisation" button and then go to "Optimise Acquiring". The fluorescence Target Sample Range has to be between 5 and 10 FI for each channel. Also select the option "Perform Optimisation Before 1st Acquisition".
	HEX	Yellow	
	ROX	Orange	
	Cy5	Red	
Exicycler™ 96 BIONEER	FAM	FAM	
	HEX	JOE	
	ROX	ROX	
	Cy5	Cy5	

Attached III: Optical measurement exposure setting

The exposure values of some thermocyclers must be adjusted for proper operation. Set exposure values as follows:

Thermocycler	Vitassay channel	Exposure values
DTlite Real-Time PCR System (DNA-Technology)	FAM	500
	HEX	500
	ROX	500
	Cy5	500
DTprime Real-time Detection Thermal Cycler (DNA-Technology)	FAM	500
	HEX	1000
	ROX	1000
	Cy5	1000

Bibliography/Bibliografía

1. Gautam R, Esona MD, Mijatovic-Rustempasic S, Ian Tam K, Gentsch JR, Bowen MD. Real-time RT-PCR assays to differentiate wild-type group A rotavirus strains from Rotarix® and RotaTeq® vaccine strains in stool samples. *Hum Vaccin Immunother* 2014; 10(3): 767-77.
2. Mijatovic-Rustempasic S, Tam KI, Kerin TK, Lewis JM, Gautam R, Quaye O, Gentsch JR, Bowen MD. Sensitive and specific quantitative detection of rotavirus A by one-step real-time reverse transcription-PCR assay without antecedent double-stranded-RNA denaturation. *J Clin Microbiol* 2013; 51(9): 3047-3054.
3. López L P, Carolina Cáceres D, López Medina E. Enfermedad por rotavirus. Características epidemiológicas, clínicas, prevención y manejo. *CCAP* 6(2): 45-55.

Trademarks

CFX™ and IQ5™ are registered trademarks of Bio-Rad Laboratories.

ABI®, QuantStudio™, StepOnePlus™ and ViiA™ are registered trademarks of Thermo Fisher Scientific Inc.

LightCycler® is a registered trademark of Roche.

Mx3000P™ and Mx3029™ are registered trademarks of Agilent Technologies.

Mastercycler™ is a registered trademark of Eppendorf.

Rotor-Gene® Q is a registered trademark of Qiagen.

SmartCycler® is a registered trademark of Cepheid.

Symbols for IVD components and reagents/ Símbolos utilizados para componentes y reactivos IVD

IVD	Producto para diagnóstico <i>in vitro</i> For in vitro diagnostic use only		Almacenar en lugar seco Keep dry
	Consultar las instrucciones de uso Consult instructions for use		Limitación de temperatura Temperature limitation
	Fecha de caducidad Use by		Fabricante Manufacturer
LOT	Número de lote Lot number		Contiene <n> test Contains sufficient for <n> test
DIL	Diluyente de muestra Buffer (sample diluent)		Número de referencia Catalogue number



Vitassay

Vitassay Healthcare S.L.U. Parque Tecnológico Walqa
Ctra. N-330 Km. 566 • 22197 Huesca (Spain)

www.vitassay.com